

# SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2209  
CALIBRATION DATE: 26-Jan-11

SBE4 CONDUCTIVITY CALIBRATION DATA  
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

## GHIJ COEFFICIENTS

g = -1.03673708e+001  
h = 1.40356545e+000  
i = 8.21131847e-004  
j = 2.08765706e-005  
CPcor = -9.5700e-008 (nominal)  
CTcor = 3.2500e-006 (nominal)

## ABCDM COEFFICIENTS

a = 7.88462876e-004  
b = 1.40356246e+000  
c = -1.03672329e+001  
d = -8.44044522e-005  
m = 3.1  
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.71550	0.00000	0.00000
-1.0000	34.7656	2.80087	5.21992	2.80087	-0.00000
0.9999	34.7659	2.97206	5.33486	2.97206	0.00001
15.0000	34.7655	4.26604	6.13401	4.26603	-0.00001
18.5000	34.7650	4.61230	6.33067	4.61230	0.00000
29.0000	34.7637	5.69467	6.90915	5.69468	0.00001
32.5001	34.7567	6.06678	7.09706	6.06678	-0.00001

Conductivity =  $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$  Siemens/meter

Conductivity =  $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$  Siemens/meter

t = temperature[°C]; p = pressure[decibars];  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

